

**Konformitätserklärung  
Declaration of Conformity**

Wir

We

**B. Braun Melsungen AG  
Carl-Braun-Str. 1  
34212 Melsungen  
Deutschland/Germany  
SRN DE-MF-000000201**erklären in eigener Verantwortung,  
dass das Produkt / die Produkte**Spritzenkappe NRFit®**Verschlusskappe  
(Artikelnummern und Basic UDI-DI siehe Anlage I)mit den Anforderungen der Medizinprodukte  
Verordnung (EU) 2017/745 übereinstimmen**Konformitätsbewertungsverfahren  
nach Anhang IX****Klassifizierung**  
gemäß Anhang VIII der oben genannten  
Verordnung  
Klasse I steril**Benannte Stelle**  
TÜV SÜD Product Service GmbH  
Kennnummer 0123**Gültig bis 2026-12-14**  
gemäß gültigem EU-Zertifikat  
(Nr. G11 012974 0626)hereby declare in our sole responsibility  
that the product/s**Syringe cap NRFit®**Closure cap  
(article numbers and Basic UDI-DI see attachment I)are in conformity with the requirements of the  
Medical Device Regulation (EU) 2017/745**Conformity Assessment Procedure  
according to annex IX****Classification**  
according to annex VIII of the Regulation named  
above  
Class I sterile**Notified Body**  
TÜV SÜD Product Service GmbH  
Identification number 0123**Valid until 2026-12-14**  
according to our valid EU Certificate  
(No. G11 012974 0626)

**Anlage I / Attachment I**

**Basic UDI-DI: 4039239000023902P**

<b>Art.-Nr. / Art. No.</b>	<b>Produktname / Product name</b>
4517501N	Syringe cap NRFit®

<b>Klasse / Class</b>
Is (sterile)

#### Document amendment information

Version	Description of the changes
1.0	Initial Version under 2017/745 MDR

Title: Declaration of Conformity - 196-012-MDR - Closure cap NRFit Initiator: Sandra ? Staufenberg

This document is signed electronically in compliance with the B. Braun electronic signature policies and procedures by following persons:

UserName: Staufenberg, Sandra (stausade)  
Title: Administrator Regulatory Affairs CoE Infusion & Pain Therapy  
Date: Thursday, 27 July 2023, 11:53 W. Europe Daylight Time  
Meaning: Document signed as Author  
=====

UserName: Brand, Thomas (brantode)  
Title: HC-QM-DE08 Vice President QM for non-active Medical Devices  
Date: Thursday, 17 August 2023, 14:15 W. Europe Daylight Time  
Meaning: Approve Document  
=====

UserName: Seidel, Stefan (seidstde)  
Title: Head of Regulatory Affairs CoE Infusion & Pain Therapy  
Date: Thursday, 17 August 2023, 14:36 W. Europe Daylight Time  
Meaning: Approve Document  
=====